




CALIBRATION CERTIFICATE SENSOREX OY

Sensorex Transmitters Type SX-912
0-2000ppm C2H4 = Isign 4-20mA, Lin

Omega Measuring Instrument Co Ltd.

Ser. no	0	500	1000	1500	2000	ppm
23918	4,0	8,0	12,0	16,0	20,0	mA



Naantali, FINLAND 21.07.2016

Esa Miettinen



~~Example~~
Example