**จัดจำหน่ายโดย** บริษัท โอเมก้า เมชเชอริ่ง อินสทรูเม้นท์ จำกัด 50/23 หมู่ 3 ด.มหาสวัสดิ์ อ.บางกรวย จ.นนทบรี 11130 Tel : 02 105 4676 Fax : 02 903 0080 ext. 6867 Email : info@omi.co.th





# Thermal Flow Sensors TA20 05/04

Probes · Probe Extensions · Probe Guides



Flow · Flow rate also combined with Temperature · Pressure

Specifications Designs Information for the user



# The principle of measurement

is based on heat transport from an electrically heated body in ambient air dependent on the relative velocity between both. Höntzsch TA sensors consist of an electrically heated Pt100 and two Pt100 temperature probes. One of the two temperature probes measures the temperature of the electrically heated Pt100, the other the temperature of the flowing air.

The feed circuit for the heated element and the temperature difference between heated element and medium are the dimensions for the flow velocity of the air.

Thermal flow sensors TA from Höntzsch are able to measure not only the lowest but also high velocities and meet all requirements for industrial application:

- · chemical corrosion resistance
- robust mechanical design
- small dimensions
- interchangeability (unrestricted allocation to the evaluation unit)
- easy to clean
- wide measuring range



Heating current I dependent on the standard velocity Nv and the temperature t

The diagram shows that the correspondence between heating current and the standard velocity is also temperature-dependent. Höntzsch flow evaluation unit VT-TA compensates this dependency over the entire temperature application range provided for one sensor respectively.

The  $T_{95}$  <u>time constant</u> of all TA sensors at air velocities of around 5 Nm/s is approx. 10 s. The time constant decreases at higher velocities. The sensors are realized in a modified Pt100 4-wire technique. This allows for maximum precision in the case of measurement of temperature and also results in the connection cable between electronic evaluation unit and sensor being able to measure up to several hundred meters. A characteristic feature to be emphasized when measuring with thermal flow sensors is that the measurement accuracy increases in decreasing flow velocities. In contrast to this, the measurement accuracy increases in increasing flow velocities in the case of vane wheel and vortex flow sensors.

#### flow measuring technology

## Measurable variables

are the standard flow velocity Nv and the temperature t. The standard volume flow and mass flow can be determined with Nv independent of pressure and temperature.

Standard flow velocity and actual flow

The standard velocity Nv is the velocity at which a gas moves when gas temperature and gas pressure are identical with the standard condiBv is the velocity at which a microscopically small particle floats in a flowing medium.

Thermal anemometers measure the standard velocity Nv, that is a velocity which is referred to so-called standard conditions. Measuring the standard velocity with a thermal flow sensor is independent of pressure and temperature.

The standard conditions for Höntzsch are +21 °C (+70 °F) and 1014 hPa (14.7 psia).

The decision whether to apply the standard or actual velocities depends entirely on the measurement requirements. The standard velocity can be converted to an actual velocity and vice versa using the universal gas equation.

Höntzsch flow evaluation units VT-TA and PTW can carry out this conversion automatically.



### Measuring tubes TA Di...

for direct, independent of position, stationary installation in pipelines with inside diameters from 10 mm ... 65 mm.Tube connection by flange, cutting ring fitting, pipe fitting, connecting nipple, etc.

TA·Di 32 GE 60 Nm/s /100/p12 for installation between flanges





## Probe designs





Model designations of thermal flow sensors TA



# Types of sensor

- TA with constant heating current at variable temperature difference between heated element and medium: normal operating mode for TA20 sensors and measuring tubes TA·Di...
- temperature sensor Pt100

#### Sensor TA20



# U199\_TA20\_D\_e\_040527

#### Sensor materials

materials coming into contact with the medium

- TA20 for 100 °C
- E stainless steel 1.4571, epoxy resin, VITON®

#### TA20 for 240 °C

E stainless steel 1.4571, VITON<sup>®</sup>, glass

#### Medium

Thermal sensors TA are calibrated with air and are primarily intended for measurements in air. Gases other than air can have varied cooling effects and can therefore influence the measurement. Disregarding this, thermal flow sensors are also suitable for measurements in singlephase gaseous flows. Measurements in gas mixtures with air or other gases are affected to such an extent that the heat transfer property of the gas mixture differentiates from the heat transfer property of the air.

The use of a TA sensor in particlecharged flows is permissible as long as the finish of the sensor material stainless steel does not change and deposits are not allowed to collect on the sensor. Mediums may even contain fibres.

Deposits or abrasions on the sensor can change the heat transmissions and so influence the measurement. The effect of condensation on the sensor is to be regarded in the same way. <u>Clean soiled sensor heads!</u>

Drifts resulting from variable <u>humidity</u> of the air are covered by our specifications for measuring uncertainty.

#### Measuring range

The smallest measurable value with velocity measurements of air is approx. 0.08 m/s, the terminal value 40 m/s or 60 m/s, air at 21 °C, 1014 hPa. Velocities of less than 0.08 m/s can be measured and are also output on the appropriate evaluation units. Velocities of less than 0.1 m/s are however not controlled by calibration.

The <u>velocity overload capacity</u> of the TA20 sensors up to 100 m/s. If the velocity in blower stream is higher than the specified nominal flow then measured values greater than the nominal flow will be issued. These values however are no longer covered for by the standard data relevant to measuring uncertainty. In the case of velocities in blower stream which are distinctly higher than the specified nominal flow, then overflow is displayed.

The <u>fatigue strength</u> for infinite time of the thermal flow sensor TA is independent of the flow conditions within the measuring range and independent of the velocity overload.



flow measuring technology

#### Working temperature range

100 -10 °C ... +100 °C **C** 240 -10 °C ... +240 °C **C C** = continuous operation

The working temperature range indication is equivalent to the temperature measuring range. Measurements of around 0 °C and less are only then possible if no ice, hair-frost, snow or other agglomerations are allowed to collect on the sensor heated element. The zero point is influenced by absolute pressures which are distinctly higher than 1014 hPa. In the case of non-compensated pressure dependence the following approximate values arise:

0 Nm/s, 5065 hPa abs/ $\approx$  5 bar: measured value 0.25 Nm/s 0 Nm/s, 10130 hPa abs/ $\approx$  10 bar: measured value 0.50 Nm/s

This influence is no longer effective at velocities higher than approx. 1 Nm/s. The standard software TA of the flow evaluation units VT-TA is equipped, since 6/97 or in the case of the PTW since 12/97, with an inquiry to the mean working pressure for the purpose of the best possible compensation of the described pressure dependence.



Sensor TA20



Sensor TA20... ZG9a

#### Measuring uncertainty

see data sheet Flow Velocity Calibration TA (U194\_EN)

#### **Connection cable**

As a rule the connection cable for probes with direct connection cable outlet is 8-wire, 2 m long and for probes with a working temperature range of up to +100 °C and +240 °C TEFLON coated. It has a continuous working temperature range of up to +260 °C, short-time up to +300 °C, and has an overall shielding for reasons of electromagnetic resistance.

When ordering please name the accompanying evaluation unit so that appropriate connector plug / connection identification can be supplied.

# Maximum working pressure, pressure dependence

- p0 sensor is not tight, not pressure resistant
- p... sensor is tight and structurally designed for pressures above atmospheric of up to ... bar (1 bar = 100 kPa)

Neither liquids nor corrosive gases must be allowed to penetrate the sensor from the cable connection side. In this respect please enquire about protected sensors before ordering.

Measurements with absolute pressure of less than approx. 0.7 bar absolute are no longer covered for by the standard data relevant to measuring uncertainty.

## Probe extensions

are used for deeper insertion of probes in mediums, when the standard probe length is not sufficient for the required insertion depth.

Extendable probes TA20 have, for this purpose, a connection thread onto which extension tubes can be fitted which are also suited for feedthrough of sensor connection cable and connector plug. In addition they are protected against mediums water or corrosive gases - which should not be allowed to penetrate the sensor from the cable connection

## Probe guide pieces SF

are used for inserting and retracting probes in and out of pipelines and ducts. The fixing device of the probe tube is to be chosen taking the temperature and pressure operating conditions into consideration.

- SFB SF with clamping <u>b</u>ush for unlimited repeated positioning of a probe at low subatmospheric and above atmospheric pressures
- SFS SF with cutting ring fitting for single/one-off positioning of a probe at higher subatmospheric and above atmospheric pressures and higher temperatures

side. Probe tubes are suitable for use with probe guide pieces, even in corrosive mediums or at high temperatures and offer a mechanically steady probe support.

#### Seal material

on the screw fittings: VITON<sup>®</sup> O-rings. VITON<sup>®</sup> is an extremely versatile corrosion resistant material and is temperature resistant in a range from -20 °C ... +240 °C **C** -25 °C ... +240 °C **S** 

**C** = continuous operation

**S** = short-time operation

The mechanical burden of the tubes on the connection threads (burden due to weight of probe and extension tube as well as force of the flowing medium) limits the number of extension tubes which can be screwed together. Besides this, the sensor must not be allowed to undergo any vibration.

For the most cases we recommend that not more than 4 extension tubes of 400 mm in length or rather 2 extension tubes of 1000 mm in length should be screwed together without additional support.



TA20...ZG8a, probe guide piece with ball valve and chain safety device

flow measuring technology

#### Coefficient / Profile factor (PF = BW)

In larger free jet as well as in larger ducts and measuring tubes the local velocity  $v_p$  will be displayed with PF = 1.000. PF is also used to calculate the local velocity  $v_p$  to the average velocity  $v_m$  in a measurement crosssection:

$$v_m = v_p \cdot PF$$

This relation applies likewise to standard and actual flow velocities.

When measuring with thermal flow sensors TA in circular measurement cross-sections with interior diameter Di then in turbulent flow the coefficient 0.84 is to be taken as a basis as long as the following conditions are adhered to

- sensor to be so positioned that the sensor heated element is centric
- sensor to be so positioned that the medium temperature sensor is within the tube flow and a constant distribution of temperature above the meas urement crosssection exists
  irrotational flow

• developed flow profile, i.e. measurement cross-section so chosen that 20 Di straight, un hindered input section, amounts to 10 Di straight, unhindered out put section.



#### **Pressure drop**

Pressure drop of several Höntzsch flow sensors when measuring in air under normal atmospheric conditions, +20 °C



#### Höntzsch GmbH

P.O. Box 1324 D-71303 Waiblingen Gottlieb-Daimler-Straße 37 D-71334 Waiblingen (Hegnach) Telephone 07151/1716-0 Facsimile 07151/58402

#### จัดจำหน่ายโดย

**บริษัท โอเมก้า เมชเชอริ่ง อินสทรูเม้นท์ จำกัด** 50/23 หมู่ 3 ด.มหาสวัสดิ์ อ.บางกรวย จ.นนทบรี 11130 Tel : 02 105 4676 Fax : 02 903 0080 ext. 6867 Email : info@omi.co.th